

Test Certificate

Test Cytotoxicity, L 929-Proliferation
DIN EN ISO 10993-5, -12, ISO 9363-1, LM SOP 4-06-01

Test material CoCrW alloy Remanium Star, lot 10973

Manufacturer Dentaurum J.P. Winkelstroeter KG, Turnstrasse 31, D-75228 Ispringen

Test number 022450-20 (the complete report is available at the client)

Test result The present results indicate that the CoCrW alloy Remanium Star does not release cytotoxic substances.



Dr. Madlon Timme
Study Director



Dipl.-Ing. (FH) Anke Warschkow
Technical Manager