



AMANNGIRRBACH

IMPLANT PROSTHETICS

 **ceramill**[®]



MPS

Multi Platform Systems
by

 **MEDENTIKA**[®]
A Straumann Group Brand

4	B-Series	compatible with	Bredent Medical® SKY®
10	BS-Series	compatible with	BEGO Implant Systems® Semados® S-Line Semados® RS/R SX-Line Semados® RI-Line
14	C-Series	compatible with	Altatec® Camlog Screw-Line - Root-Line 2®
20	CX-Series	compatible with	Medentis Medical® ICX®
24	D-Series	compatible with	Altatec® Conelog®
30	E-Series	compatible with	Nobel Biocare® NobelReplace® Tapered
36	EV-Series	compatible with	Dentsply® Implants ASTRA TECH OsseoSpeed® EV
42	F-Series	compatible with	Nobel Biocare® Nobel Active® NobelReplace® Tapered
48	H-Series	compatible with	Biomet 3i® Certain®
54	I-Series	compatible with	Biomet 3i® External Hex®
60	K-Series	compatible with	Nobel Biocare® Brånemark®
66	L-Series	compatible with	Straumann® Bone Level®
72	N-Series	compatible with	Straumann® Tissue Level®
78	R-Series	compatible with	Zimmer Dental® MIS® BioHorizons®
86	S-Series	compatible with	DENTSPLY Implants® ASTRA TECH OsseoSpeed® TX
92	T-Series	compatible with	DENTSPLY Implants® XiVE® S
98	Y-Series	compatible with	DENTSPLY Implants® ANKYLOS®
104	Tools		

» B-SERIES «




The B-Series abutments are compatible with



Bredent Medical® SKY®

DETAILED VIEW

Exact view of the direction of the angulation (indexing) of all angled abutments.



Type 1

angled over flat

Laboratory implant

- Titanium grade 5 KV



Implant connection

D 3,5/4,0

Article No.

B 50

Abutment screw

- Titanium grade 5 KV
- Torx T6



Article No.

B 60

Please note: The correct choice of the abutment screw is stated with the respective abutment in the catalogue.

Laboratory screw

- red coated



Article No.

B 60 L

Please note: The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm



Implant connection

D 3,5

Chimney height

3.5 mm

Gingiva height

0,3 mm

Article No.

B 1000

Abutment screw

B 60

Laboratory screw

B 60 L

Please note: This is used with Scanbody 2. generation.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm



Implant connection	D 3,5
Chimney height	5.5 mm
Gingiva height	0,3 mm
Article No.	B 1100
Abutment screw	B 60
Laboratory screw	B 60 L

Please note: This is used with Scanbody 2. generation.

Titanium base Bridges/ bars 2nd Generation

- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm



Implant connection	D 3,5
Chimney height	3.5 mm
Gingiva height	0,3 mm
Article No.	B 1500
Abutment screw	B 60
Laboratory screw	B 60 L

Please note: This is used with Scanbody 2. generation.

TI-Forms®-Abutment for Ceramill

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm



Implant connection	D 3,5
Diameter	11.5 mm
Article No.	B 9000-AG
Abutment screw	B 60
Laboratory screw	B 60 L

Please note: This is used with Scanbody 2. generation.

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection	D 3,5/4,0
Article No.	B 1400

Please note: The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 25 Ncm



Implant connection	D 4,0	D 4,0	D 4,0	D 4,0	D 4,0
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	B 4000	B 4100	B 4200	B 4300	B 4400

MedentiBASE® prosthetic components

- B 4840 MedentiBASE® titanium base incl. bridge screw
- B 7600 MedentiBASE® bridge screw Torx T6
- Recommended torque: 15 Ncm



Material	Titanium grade 5 KV	Titanium grade 5 KV
Article No.	B 4840	B 7600

MedentiBASE® adhesive cap for passive-fit

- B 4800 MedentiBASE® adhesive cap Short incl. bridge screw
- B 4820 MedentiBASE® adhesive cap Long incl. bridge screw



Version	Short	Long
Article No.	B 4800	B 4820

MedentiBASE® scanbody

- Titanium grade 5 KV specially coated
- incl. bridge screw



Article No.	B 4900
-------------	--------

Please note: The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® accessories

- M 11-6 Placement instrument MedentiBASE® abutment



Article No.	M 11-6
-------------	--------

» BS-SERIES «



The BS-Series abutments are compatible with



BEGO Implant Systems®
Semados® S-Line
Semados® RS/RSX-Line
Semados® RI-Line

BS

Series

Abutment screw

- Titanium grade 5 KV
- Hex 1.26



Article No.

BS 60

Please note: The correct choice of the abutment screw is stated with the respective abutment in the catalogue.

Laboratory screw

- red coated



Article No.

BS 60 L

Please note: The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 30 Ncm



	D 4,1	D 4,5	D 5,5
Implant connection	D 4,1	D 4,5	D 5,5
Chimney height	3.5 mm	3.5 mm	3.5 mm
Gingiva height	0,1 mm	0,1 mm	0,1 mm
Article No.	BS 1010	BS 1020	BS 1030
Abutment screw	BS 60	BS 60	BS 60
Laboratory screw	BS 60 L	BS 60 L	BS 60 L

Please note: This is used with Scanbody 2. generation.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 30 Ncm



	D 4,1	D 4,5	D 5,5
Implant connection	D 4,1	D 4,5	D 5,5
Chimney height	5.5 mm	5.5 mm	5.5 mm
Gingiva height	0,1 mm	0,1 mm	0,1 mm
Article No.	BS 1110	BS 1120	BS 1130
Abutment screw	BS 60	BS 60	BS 60
Laboratory screw	BS 60 L	BS 60 L	BS 60 L

Please note: This is used with Scanbody 2. generation.

Titanium base Bridges/ bars 2nd Generation

- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 30 Ncm



Implant connection	D 4,1	D 4,5
Chimney height	3.5 mm	3.5 mm
Gingiva height	0,15 mm	0,15 mm
Article No.	BS 1510	BS 1520
Abutment screw	BS 60	BS 60
Laboratory screw	BS 60 L	BS 60 L

Please note: This is used with Scanbody 2. generation.

TI-Forms®-Abutment for Ceramill

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 30 Ncm



Implant connection	D 3,25-5,5
Diameter	11.5 mm
Article No.	BS 9000-AG
Abutment screw	BS 60
Laboratory screw	BS 60 L

Please note: This is used with Scanbody 2. generation.

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection	D 3,25-5,5
Article No.	BS 1400

Please note: The Scanbody is sterilisable and for intra-oral scanning.

» C-SERIES «



The C-Series abutments are compatible with




Altatec®
Camlog Screw-Line – Root-Line 2®

C
Series

Laboratory implant

- Titanium grade 5 KV



Implant connection	D 3,3	D 3,8	D 4,3	D 5,0
Article No.	C 50	C 51	C 52	C 53

Abutment screw

- Titanium grade 5 KV
- Hex 1.26



Article No.	C 60	C 61	C 62
-------------	------	------	------

Please note: The correct choice of the abutment screw is stated with the respective abutment in the catalogue.

Laboratory screw

- red coated



Article No.	C 61 L	C 62 L
-------------	--------	--------

Please note: The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 20 Ncm



Implant connection	D 3,3	D 3,8	D 4,3	D 5,0	D 6,0
Chimney height	3.5 mm	3.5 mm	3.5 mm	3.5 mm	3.5 mm
Gingiva height	0,3 mm	0,3 mm	0,3 mm	0,3 mm	0,3 mm
Article No.	C 1000	C 1010	C 1020	C 1030	C 1040
Abutment screw	C 61	C 61	C 61	C 62	C 62
Laboratory screw	C 61 L	C 61 L	C 61 L	C 62 L	C 62 L

Please note: This is used with Scanbody 2. generation.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 20 Ncm



Implant connection	D 3,3	D 3,8	D 4,3	D 5,0	D 6,0
Chimney height	5,5 mm	5,5 mm	5,5 mm	5,5 mm	5,5 mm
Gingiva height	0,3 mm	0,3 mm	0,3 mm	0,3 mm	0,3 mm
Article No.	C 1100	C 1110	C 1120	C 1130	C 1140
Abutment screw	C 61	C 61	C 61	C 62	C 62
Laboratory screw	C 61 L	C 61 L	C 61 L	C 62 L	C 62 L

Please note: This is used with Scanbody 2. generation.

Titanium base Bridges/ bars 2nd Generation

- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 20 Ncm



Implant connection	D 3,3	D 3,8	D 4,3
Chimney height	3,5 mm	3,5 mm	3,5 mm
Gingiva height	0,3 mm	0,3 mm	0,3 mm
Article No.	C 1500	C 1510	C 1520
Abutment	C 60	C 60	C 60

Please note: This is used with Scanbody 2. generation.

TI-Forms®-Abutment for Ceramill

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 20 Ncm



Implant connection	D 3,3	D 3,8	D 4,3	D 5,0
Diameter	11,5 mm	11,5 mm	11,5 mm	11,5 mm
Article No.	C 9000-AG	C 9010-AG	C 9020-AG	C 9030-AG
Abutment screw	C 61	C 61	C 61	C 62
Laboratory screw	C 61 L	C 61 L	C 61 L	C 62 L

Please note: This is used with Scanbody 2. generation.

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection	D 3,3	D 3,8	D 4,3	D 5,0/6,0
Article No.	C 1400	C 1410	C 1420	C 1430

Please note: The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 20 Ncm



Implant connection	D 3,3	D 3,3	D 3,3	D 3,3	D 3,3
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	C 4000	C 4100	C 4200	C 4300	C 4400

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 30 Ncm



Implant connection	D 3,8	D 3,8	D 3,8	D 3,8	D 3,8
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	C 4010	C 4110	C 4210	C 4310	C 4410

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 30 Ncm



Implant connection	D 4,3	D 4,3	D 4,3	D 4,3	D 4,3
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	C 4020	C 4120	C 4220	C 4320	C 4420

MedentiBASE® prosthetic components

- C 4840 MedentiBASE® titanium base incl. bridge screw
- C 7600 MedentiBASE® bridge screw Hex 1.26
- Recommended torque: 15 Ncm



Material	Titanium grade 5 KV	Titanium grade 5 KV
Article No.	C 4840	C 7600

MedentiBASE® adhesive cap for passive-fit

- C 4800 MedentiBASE® adhesive cap Short incl. bridge screw
- C 4820 MedentiBASE® adhesive cap Long incl. bridge screw



Version	Short	Long
Article No.	C 4800	C 4820

MedentiBASE® scanbody

- Titanium grade 5 KV specially coated
- incl. bridge screw



Article No.

C 4900

Please note: The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® accessories

- M 11-6 Placement instrument
MedentiBASE® abutment



Article No.

M 11-6

»» CX-SERIES ««



The CX-Series abutments are compatible with



Medentis Medical®
ICX®

CX
Series

Abutment screw

- Titanium grade 5 KV
- Hex 1.4


Article No.

CX 60

Please note:

The correct choice of the abutment screw is stated with the respective abutment in the catalogue.

Laboratory screw

- red coated


Article No.

CX 60 L

Please note:

The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 30 Ncm



Implant connection	D 3,75-4,8	D 3,75-4,8
Chimney height	3.5 mm	3.5 mm
Gingiva height	0,65 mm	1,15 mm
Article No.	CX 1000	CX 1200
Abutment screw	CX 60	CX 60
Laboratory screw	CX 60 L	CX 60 L

Please note:

This is used with Scanbody 2. generation.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 30 Ncm



Implant connection	D 3,75-4,8	D 3,75-4,8
Chimney height	5.5 mm	5.5 mm
Gingiva height	0,65 mm	1,15 mm
Article No.	CX 1100	CX 1300
Abutment screw	CX 60	CX 60
Laboratory screw	CX 60 L	CX 60 L

Please note:

This is used with Scanbody 2. generation.

Titanium base Bridges/ bars 2nd Generation

- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 30 Ncm



Implant connection	D 3,75-4,8
Chimney height	3.5 mm
Gingiva height	0,65 mm
Article No.	CX 1500
Abutment screw	CX 60
Laboratory screw	CX 60 L

Please note: This is used with Scanbody 2. generation.

TI-Forms®-Abutment for Ceramill

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 30 Ncm



Implant connection	D 3,75-4,8
Diameter	11.5 mm
Article No.	CX 9000-AG
Abutment screw	CX 60
Laboratory screw	CX 60 L

Please note: This is used with Scanbody 2. generation.

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection	D 3,75-4,8
Article No.	CX 1400

Please note: The Scanbody is sterilisable and for intra-oral scanning.

»» D-SERIES ««



The D-Series abutments are compatible with




Altatec®
Conelog®

D
Series


DETAILED VIEW

Exact view of the direction of the angulation (indexing) of all angled abutments.



Type 1

angled over flat




Type 2

angled over cam

Laboratory implant

- Stainless steel


Implant connection

D 3,3

D 3,8/4,3

Article No.

D 50

D 51

Abutment screw

- Titanium grade 5 KV
- Hex 1.26


Article No.

D 60

Please note:

The correct choice of the abutment screw is stated with the respective abutment in the catalogue.

Laboratory screw

- red coated


Article No.

D 60 L

Please note:

The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 20 Ncm


Implant connection

D 3,3

D 3,3

D 3,8/4,3

D 3,8/4,3

Chimney height

3,5 mm

3,5 mm

3,5 mm

3,5 mm

Gingiva height

0,65 mm

1,15 mm

0,65 mm

1,15 mm

Article No.

D 1000

D 1200

D 1010

D 1210

Abutment screw

D 60

D 60

D 60

D 60

Laboratory screw

D 60 L

D 60 L

D 60 L

D 60 L

Please note:

This is used with Scanbody 2. generation.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 20 Ncm



Implant connection	D 3,3	D 3,3	D 3,8/4,3	D 3,8/4,3
Chimney height	5.5 mm	5.5 mm	5.5 mm	5.5 mm
Gingiva height	0,65 mm	1,15 mm	0,65 mm	1,15 mm
Article No.	D 1100	D 1300	D 1110	D 1310
Abutment screw	D 60	D 60	D 60	D 60
Laboratory screw	D 60 L	D 60 L	D 60 L	D 60 L

Please note: This is used with Scanbody 2. generation.

Titanium base Bridges/ bars 2nd Generation

- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 20 Ncm



Implant connection		D 3,3	D 3,8/4,3
Chimney height		3.5 mm	3.5 mm
Gingiva height		0,65 mm	0,65 mm
Article No.		D 1500	D 1510
Abutment screw		D 60	D 60
Laboratory screw		D 60 L	D 60 L

Please note: This is used with Scanbody 2. generation.

TI-Forms®-Abutment for Ceramill

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 20 Ncm



Implant connection		D 3,3	D 3,8/4,3
Diameter		11.5 mm	11.5 mm
Article No.		D 9000-AG	D 9010-AG
Abutment screw		D 60	D 60
Laboratory screw		D 60 L	D 60 L

Please note: This is used with Scanbody 2. generation.

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection		D 3,3	D 3,8/4,3
Article No.		D 1400	D 1410

Please note: The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 20 Ncm



Implant connection	D 3,3	D 3,3	D 3,3	D 3,3	D 3,3
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	D 4000	D 4100	D 4200	D 4300	D 4400

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 20 Ncm



Implant connection	D 3,8/4,3	D 3,8/4,3	D 3,8/4,3	D 3,8/4,3	D 3,8/4,3
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	D 4010	D 4110	D 4210	D 4310	D 4410

MedentiBASE® prosthetic components

- D 4840 MedentiBASE® titanium base incl. bridge screw
- D 7600 MedentiBASE® bridge screw Hex 1.26
- Recommended torque: 15 Ncm



Material	Titanium grade 5 KV	Titanium grade 5 KV
Article No.	D 4840	D 7600

MedentiBASE® adhesive cap for passive-fit

- D 4800 MedentiBASE® adhesive cap Short incl. bridge screw
- D 4820 MedentiBASE® adhesive cap Long incl. bridge screw



Version	Short	Long
Article No.	D 4800	D 4820

MedentiBASE® scanbody

- Titanium grade 5 KV specially coated
- incl. bridge screw



Article No.	D 4900
-------------	--------

Please note: The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® accessories

- M 11-6 Placement instrument
MedentiBASE® abutment



Article No.

M 11-6

» E-SERIES «



The E-Series abutments are compatible with




Nobel Biocare NobelReplace® Tapered

E
Series


DETAILED VIEW

Exact view of the direction of the angulation (indexing) of all angled abutments.



Type 1

angled over flat




Type 2

angled over cam

Laboratory implant

- Titanium grade 5 KV



Implant connection	NP 3,5	RP 4,3	WP 5,0	D 6,0
Article No.	E 50	E 51	E 52	E 53

Abutment screw

- Titanium grade 5 KV
- UG



Article No.	E 60	E 61
-------------	------	------

Please note: The correct choice of the abutment screw is stated with the respective abutment in the catalogue.

Laboratory screw

- red coated



Article No.	E 60 L	E 61 L
-------------	--------	--------

Please note: The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm




Implant connection	NP 3,5	RP 4,3	WP 5,0	D 6,0
Chimney height	3.5 mm	3.5 mm	3.5 mm	3.5 mm
Gingiva height	0,3 mm	0,3 mm	0,3 mm	0,3 mm
Article No.	E 1000	E 1010	E 1020	E 1030
Abutment screw	E 60	E 61	E 61	E 61
Laboratory screw	E 60 L	E 61 L	E 61 L	E 61 L

Please note: This is used with Scanbody 2. generation.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection	NP 3,5	RP 4,3	WP 5,0	D 6,0
Chimney height	5.5 mm	5.5 mm	5.5 mm	5.5 mm
Gingiva height	0,3 mm	0,3 mm	0,3 mm	0,3 mm
Article No.	E 1100	E 1110	E 1120	E 1130
Abutment screw	E 60	E 61	E 61	E 61
Laboratory screw	E 60 L	E 61 L	E 61 L	E 61 L

Please note: This is used with Scanbody 2. generation.

Titanium base Bridges/ bars 2nd Generation

- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection	NP 3,5	RP 4,3	WP 5,0
Chimney height	3.5 mm	3.5 mm	3.5 mm
Gingiva height	0,3 mm	0,3 mm	0,3 mm
Article No.	E 1500	E 1510	E 1520
Abutment screw	E 60	E 61	E 61
Laboratory screw	E 60 L	E 61 L	E 61 L

Please note: This is used with Scanbody 2. generation.

TI-Forms®-Abutment for Ceramill

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection	NP 3,5	RP 4,3	WP 5,0
Diameter	11.5 mm	11.5 mm	11.5 mm
Article No.	E 9000-AG	E 9010-AG	E 9020-AG
Abutment screw	E 60	E 61	E 61
Laboratory screw	E 60 L	E 61 L	E 61 L

Please note: This is used with Scanbody 2. generation.

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection	NP 3,5	RP 4,3	WP 5,0	D 6,0
Article No.	E 1400	E 1410	E 1420	E 1430

Please note: The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 35 Ncm



Implant connection	NP 3,5	NP 3,5	NP 3,5	NP 3,5	NP 3,5
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	E 4000	E 4100	E 4200	E 4300	E 4400

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 35 Ncm



Implant connection	RP 4,3	RP 4,3	RP 4,3	RP 4,3	RP 4,3
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	E 4010	E 4110	E 4210	E 4310	E 4410

MedentiBASE® prosthetic components

- E 4840 MedentiBASE® titanium base incl. bridge screw
- E 7600 MedentiBASE® bridge screw UG
- Recommended torque: 15 Ncm



Material	Titanium grade 5 KV	Titanium grade 5 KV
Article No.	E 4840	E 7600

MedentiBASE® adhesive cap for passive-fit

- E 4800 MedentiBASE® adhesive cap Short incl. bridge screw
- E 4820 MedentiBASE® adhesive cap Long incl. bridge screw



Version	Short	Long
Article No.	E 4800	E 4820

MedentiBASE® scanbody

- Titanium grade 5 KV specially coated
- incl. bridge screw



Article No.	E 4900
-------------	--------

Please note: The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® accessories

- M 11-6 Placement instrument
MedentiBASE® abutment



Article No.

M 11-6

» EV-SERIES «



The EV-Series abutments are compatible with



Dentsply® Implants
ASTRA TECH OsseoSpeed® EV

EV

Series

Laboratory implant


- Titanium grade 5 KV



Implant connection	D 3,0	D 3,6	D 4,2	D 4,8	D 5,4
Article No.	EV 50	EV 51	EV 52	EV 53	EV 54

Abutment screw

- Titanium grade 5 KV
- Hex 1.26



Article No.	EV 60	EV 61	EV 62	EV 63
-------------	-------	-------	-------	-------

Please note: The correct choice of the abutment screw is stated with the respective abutment in the catalogue.

Laboratory screw

- red coated



Article No.	EV 60 L	EV 61 L	EV 62 L	EV 63 L
-------------	---------	---------	---------	---------

Please note: The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm



Implant connection	D 3,0	D 3,0	D 3,6	D 3,6	D 4,2
Chimney height	3,5 mm	3,5 mm	3,5 mm	3,5 mm	3,5 mm
Gingiva height	0,65 mm	1,15 mm	0,65 mm	1,15 mm	0,65 mm
Article No.	EV 1000	EV 1200	EV 1010	EV 1210	EV 1020
Abutment screw	EV 60	EV 60	EV 61	EV 61	EV 62
Laboratory screw	EV 60 L	EV 60 L	EV 61 L	EV 61 L	EV 62 L

Please note: This is used with Scanbody 2. generation.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm



Implant connection	D 4,2	D 4,8	D 4,8	D 5,4	D 5,4
Chimney height	3,5 mm	3,5 mm	3,5 mm	3,5 mm	3,5 mm
Gingiva height	1,15 mm	0,65 mm	1,15 mm	0,65 mm	1,15 mm
Article No.	EV 1220	EV 1030	EV 1230	EV 1040	EV 1240
Abutment screw	EV 62	EV 63	EV 63	EV 63	EV 63
Laboratory screw	EV 62 L	EV 63 L	EV 63 L	EV 63 L	EV 63 L

Please note: This is used with Scanbody 2. generation.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm



Implant connection	D 3,0	D 3,0	D 3,6	D 3,6	D 4,2
Chimney height	5,5 mm	5,5 mm	5,5 mm	5,5 mm	5,5 mm
Gingiva height	0,65 mm	1,15 mm	0,65 mm	1,15 mm	0,65 mm
Article No.	EV 1100	EV 1300	EV 1110	EV 1310	EV 1120
Abutment screw	EV 60	EV 60	EV 61	EV 61	EV 62
Laboratory screw	EV 60 L	EV 60 L	EV 61 L	EV 61 L	EV 62 L

Please note: This is used with Scanbody 2. generation.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm



Implant connection	D 4,2	D 4,8	D 4,8	D 5,4	D 5,4
Chimney height	5,5 mm	5,5 mm	5,5 mm	5,5 mm	5,5 mm
Gingiva height	1,15 mm	0,65 mm	1,15 mm	0,65 mm	1,15 mm
Article No.	EV 1320	EV 1130	EV 1330	EV 1140	EV 1340
Abutment screw	EV 62	EV 63	EV 63	EV 63	EV 63
Laboratory screw	EV 62 L	EV 63 L	EV 63 L	EV 63 L	EV 63 L

Please note: This is used with Scanbody 2. generation.

TI-Forms®-Abutment for Ceramill

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm



Implant connection	D 3,0	D 3,6	D 4,2	D 4,8	D 5,4
Diameter	11.5 mm	11.5 mm	11.5 mm	11.5 mm	11.5 mm
Article No.	EV 9000-AG	EV 9010-AG	EV 9020-AG	EV 9030-AG	EV 9040-AG
Abutment screw	EV 60	EV 61	EV 62	EV 63	EV 63
Laboratory screw	EV 60 L	EV 61 L	EV 62 L	EV 63 L	EV 63 L

Please note: This is used with Scanbody 2. generation.

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection	D 3,0	D 3,6	D 4,2	D 4,8	D 5,4
Article No.	EV 1400	EV 1410	EV 1420	EV 1430	EV 1440

Please note: The Scanbody is sterilisable and for intra-oral scanning.

» F-SERIES «




The F-Series abutments are compatible with



Nobel Biocare® NobelActive® / NobelReplace® Conical


DETAILED VIEW

Exact view of the direction of the angulation (indexing) of all angled abutments.



Type 1

angled over flat



Type 2

angled over corner

F
Series

Laboratory implant

- Titanium grade 5 KV



Implant connection

D 3,0

NP 3,5

RP 4,3/5,0

Article No.

F 52

F 50

F 51

Abutment screw

- Titanium grade 5 KV
- UG



Article No.

F 60

F 61

F 66

Please note:

The correct choice of the abutment screw is stated with the respective abutment in the catalogue.

Laboratory screw

- red coated



Article No.

F 60 L

F 61 L

F 66 L

Please note:

The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque:
15 Ncm: D 3,0
35 Ncm: NP 3,5
35 Ncm: RP 4,3/5,0



Implant connection

D 3,0

NP 3,5

NP 3,5

RP 4,3/5,0

RP 4,3/5,0

Chimney height

3,5 mm

3,5 mm

3,5 mm

3,5 mm

3,5 mm

Gingiva height

0,65 mm

0,65 mm

1,15 mm

0,65 mm

1,15 mm

Article No.

F 1020

F 1000

F 1200

F 1010

F 1210

Abutment screw

F 66

F 60

F 60

F 61

F 61

Laboratory screw

F 66 L

F 60 L

F 60 L

F 61 L

F 61 L

Please note:

This is used with Scanbody 2. generation.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque:
15 Ncm: D 3,0
35 Ncm: NP 3,5
35 Ncm: RP 4,3/5,0



Implant connection	D 3,0	NP 3,5	NP 3,5	RP 4,3/5,0	RP 4,3/5,0
Chimney height	5,5 mm	5,5 mm	5,5 mm	5,5 mm	5,5 mm
Gingiva height	0,65 mm	0,65 mm	1,15 mm	0,65 mm	1,15 mm
Article No.	F 1120	F 1100	F 1300	F 1110	F 1310
Abutment screw	F 66	F 60	F 60	F 61	F 61
Laboratory screw	F 66 L	F 60 L	F 60 L	F 61 L	F 61 L

Please note: This is used with Scanbody 2. generation.

Titanium base Bridges/ bars 2nd Generation

- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque:
15 Ncm: D 3,0
35 Ncm: NP 3,5
35 Ncm: RP 4,3/5,0




Implant connection	D 3,0	NP 3,5	RP 4,3/5,0
Chimney height	3,5 mm	3,5 mm	3,5 mm
Gingiva height	0,65 mm	0,65 mm	0,65 mm
Article No.	F 1520	F 1500	F 1510
Abutment screw	F 66	F 60	F 61
Laboratory screw	F 66 L	F 60 L	F 61 L

Please note: This is used with Scanbody 2. generation.

TI-Forms®-Abutment for Ceramill

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection	NP 3,5	RP 4,3/5,0
Diameter	11,5 mm	11,5 mm
Article No.	F 9000-AG	F 9010-AG
Abutment screw	F 60	F 61
Laboratory screw	F 60 L	F 61 L

Please note: This is used with Scanbody 2. generation.

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection	D 3,0	NP 3,5	RP 4,3/5,0
Article No.	F 1420	F 1400	F 1410

Please note: The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 35 Ncm



Implant connection	NP 3,5	NP 3,5	NP 3,5	NP 3,5	NP 3,5
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	F 4000	F 4100	F 4200	F 4300	F 4400

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 35 Ncm



Implant connection	RP 4,3/5,0	RP 4,3/5,0	RP 4,3/5,0	RP 4,3/5,0	RP 4,3/5,0
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	F 4010	F 4110	F 4210	F 4310	F 4410

MedentiBASE® prosthetic components

- F 4840 MedentiBASE® titanium base incl. bridge screw
- F 7600 MedentiBASE® bridge screw UG
- Recommended torque: 15 Ncm



Material	Titanium grade 5 KV	Titanium grade 5 KV
Article No.	F 4840	F 7600

MedentiBASE® adhesive cap for passive-fit

- F 4800 MedentiBASE® adhesive cap Short incl. bridge screw
- F 4820 MedentiBASE® adhesive cap Long incl. bridge screw



Version	Short	Long
Article No.	F 4800	F 4820

MedentiBASE® scanbody

- Titanium grade 5 KV specially coated
- incl. bridge screw



Article No.	F 4900
-------------	--------

Please note: The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® accessories

- M 11-6 Placement instrument
MedentiBASE® abutment



Article No.

M 11-6

» H-SERIES «




The H-Series abutments are compatible with



BIOMET 3i® Certain®


DETAILED VIEW

Exact view of the direction of the angulation (indexing) of all angled abutments.



Type 1

angled over flat



Type 2

angled over corner

Laboratory implant

- Titanium grade 5 KV



Implant connection

D 3,4

D 4,1

D 5,0

Article No.

H 50

H 51

H 52

Abutment screw

- Titanium grade 5 KV
- Hex 1.2



Article No.

H 60

Please note:

The correct choice of the abutment screw is stated with the respective abutment in the catalogue.

Laboratory screw

- red coated



Article No.

H 60 L

Please note:

The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 20 Ncm



Implant connection

D 3,4

D 4,1

D 5,0

Chimney height

3.5 mm

3.5 mm

3.5 mm

Gingiva height

0,3 mm

0,3 mm

0,3 mm

Article No.

H 1000

H 1010

H 1020

Abutment screw

H 60

H 60

H 60

Laboratory screw

H 60 L

H 60 L

H 60 L

Please note:

This is used with Scanbody 2. generation.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 20 Ncm



Implant connection	D 3,4	D 4,1	D 5,0
Chimney height	5.5 mm	5.5 mm	5.5 mm
Gingiva height	0,3 mm	0,3 mm	0,3 mm
Article No.	H 1100	H 1110	H 1120
Abutment screw	H 60	H 60	H 60
Laboratory screw	H 60 L	H 60 L	H 60 L

Please note: This is used with Scanbody 2. generation.

Titanium base Bridges/ bars 2nd Generation

- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 20 Ncm



Implant connection	D 3,4	D 4,1	D 5,0
Chimney height	3.5 mm	3.5 mm	3.5 mm
Gingiva height	0,3 mm	0,3 mm	0,3 mm
Article No.	H 1500	H 1510	H 1520
Abutment screw	H 60	H 60	H 60
Laboratory screw	H 60 L	H 60 L	H 60 L

Please note: This is used with Scanbody 2. generation.

TI-Forms®-Abutment for Ceramill

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 20 Ncm



Implant connection	D 3,4	D 4,1	D 5,0
Diameter	11.5 mm	11.5 mm	11.5 mm
Article No.	H 9000-AG	H 9010-AG	H 9020-AG
Abutment screw	H 60	H 60	H 60
Laboratory screw	H 60 L	H 60 L	H 60 L

Please note: This is used with Scanbody 2. generation.

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection	D 3,4	D 4,1/5,0
Article No.	H 1400	H 1410

Please note: The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 20 Ncm



Implant connection	D 3,4	D 3,4	D 3,4	D 3,4	D 3,4
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	H 4000	H 4100	H 4200	H 4300	H 4400

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 20 Ncm



Implant connection	D 4,1	D 4,1	D 4,1	D 4,1	D 4,1
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	H 4010	H 4110	H 4210	H 4310	H 4410

MedentiBASE® prosthetic components

- H 4840 MedentiBASE® titanium base incl. bridge screw
- H 7600 MedentiBASE® bridge screw Hex 1.2
- Recommended torque: 15 Ncm



Material	Titanium grade 5 KV	Titanium grade 5 KV
Article No.	H 4840	H 7600

MedentiBASE® adhesive cap for passive-fit

- H 4800 MedentiBASE® adhesive cap Short incl. bridge screw
- H 4820 MedentiBASE® adhesive cap Long incl. bridge screw



Version	Short	Long
Article No.	H 4800	H 4820

MedentiBASE® scanbody

- Titanium grade 5 KV specially coated
- incl. bridge screw



Article No.	H 4900
-------------	--------

Please note: The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® accessories

- M 11-6 Placement instrument
MedentiBASE® abutment



Article No.

M 11-6

» I-SERIES «




The I-Series abutments are compatible with



BIOMET 3i® External Hex


DETAILED VIEW

Exact view of the direction of the angulation (indexing) of all angled abutments.



Type 1

angled over flat



Type 2

angled over corner

Laboratory implant

- Titanium grade 5 KV



Implant connection

D 3,4

D 4,1

D 5,0

Article No.

I 50

I 51

I 52

Abutment screw

- Titanium grade 5 KV
- Hex 1.26



Article No.

I 61

Please note:

The correct choice of the abutment screw is stated with the respective abutment in the catalogue.

Laboratory screw

- red coated



Article No.

I 61 L

Please note:

The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection

D 3,4

D 4,1

D 5,0

Chimney height

3.5 mm

3.5 mm

3.5 mm

Gingiva height

0,6 mm

0,6 mm

0,6 mm

Article No.

I 1000

I 1010

I 1020

Abutment screw

I 61

I 61

I 61

Laboratory screw

I 61 L

I 61 L

I 61 L

Please note:

This is used with Scanbody 2. generation.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection	D 3,4	D 4,1	D 5,0
Chimney height	5.5 mm	5.5 mm	5.5 mm
Gingiva height	0,6 mm	0,6 mm	0,6 mm
Article No.	I 1100	I 1110	I 1120
Abutment screw	I 61	I 61	I 61
Laboratory screw	I 61 L	I 61 L	I 61 L

Please note: This is used with Scanbody 2. generation.

Titanium base Bridges/ bars 2nd Generation

- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection	D 3,4	D 4,1	D 5,0
Chimney height	3.5 mm	3.5 mm	3.5 mm
Gingiva height	0,6 mm	0,6 mm	0,6 mm
Article No.	I 1500	I 1510	I 1520
Abutment screw	I 61	I 61	I 61
Laboratory screw	I 61 L	I 61 L	I 61 L

Please note: This is used with Scanbody 2. generation.

TI-Forms®-Abutment for Ceramill

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection	D 3,4	RP 4,1	D 5,0
Diameter	11.5 mm	11.5 mm	11.5 mm
Article No.	I 9000-AG	I 9010-AG	I 9020-AG
Abutment screw	I 61	I 61	I 61
Laboratory screw	I 61 L	I 61 L	I 61 L

Please note: This is used with Scanbody 2. generation.

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection	D 3,4	D 4,1/5,0
Article No.	I 1400	I 1410

Please note: The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 35 Ncm



Implant connection	D 3,4	D 3,4	D 3,4
Gingiva height	2,5 mm	3,5 mm	4,5 mm
Article No.	I 4000	I 4100	I 4200

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 35 Ncm



Implant connection	D 4,1	D 4,1	D 4,1
Gingiva height	2,5 mm	3,5 mm	4,5 mm
Article No.	I 4010	I 4110	I 4210

MedentiBASE® prosthetic components

- I 4840 MedentiBASE® titanium base incl. bridge screw
- I 7600 MedentiBASE® bridge screw Hex 1.26
- Recommended torque: 15 Ncm



Material	Titanium grade 5 KV	Titanium grade 5 KV
Article No.	I 4840	I 7600

MedentiBASE® adhesive cap for passive-fit

- I 4800 MedentiBASE® adhesive cap Short incl. bridge screw
- I 4820 MedentiBASE® adhesive cap Long incl. bridge screw



Version	Short	Long
Article No.	I 4800	I 4820

MedentiBASE® scanbody

- Titanium grade 5 KV specially coated
- incl. bridge screw



Article No.	I 4900
-------------	--------

Please note: The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® accessories

- M 11-6 Placement instrument
MedentiBASE® abutment



Article No.

M 11-6

» K-SERIES «




The K-Series abutments are compatible with



Nobel Biocare® Brånemark®


DETAILED VIEW

Exact view of the direction of the angulation (indexing) of all angled abutments.



Type 1

angled over flat



Type 2

angled over corner

Laboratory implant

- Titanium grade 5 KV



Implant connection
Article No.

NP 3,5
K 50

RP 4,1
K 51

WP 5,1
K 52

Abutment screw

- Titanium grade 5 KV
- UG



Article No.

K 60

K 61

K 62

Please note: The correct choice of the abutment screw is stated with the respective abutment in the catalogue.

Laboratory screw

- red coated



Article No.

K 60 L

K 61 L

K 62 L

Please note: The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection

NP 3,5

RP 4,1

WP 5,1

Chimney height

3.5 mm

3.5 mm

3.5 mm

Gingiva height

0,6 mm

0,6 mm

0,6 mm

Article No.

K 1000

K 1010

K 1020

Abutment screw

K 60

K 61

K 62

Laboratory screw

K 60 L

K 61 L

K 62 L

Please note: This is used with Scanbody 2. generation.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection	NP 3,5	RP 4,1	WP 5,1
Chimney height	5.5 mm	5.5 mm	5.5 mm
Gingiva height	0,6 mm	0,6 mm	0,6 mm
Article No.	K 1100	K 1110	K 1120
Abutment screw	K 60	K 61	K 62
Laboratory screw	K 60 L	K 61 L	K 62 L

Please note: This is used with Scanbody 2. generation.

Titanium base Bridges/ bars 2nd Generation

- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection	NP 3,5	RP 4,1	WP 5,1
Chimney height	3.5 mm	3.5 mm	3.5 mm
Gingiva height	0,6 mm	0,6 mm	0,6 mm
Article No.	K 1500	K 1510	K 1520
Abutment screw	K 60	K 61	K 62
Laboratory screw	K 60 L	K 61 L	K 62 L

Please note: This is used with Scanbody 2. generation.

TI-Forms®-Abutment for Ceramill

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection	NP 3,5	RP 4,1	WP 5,1
Diameter	11.5 mm	11.5 mm	11.5 mm
Article No.	K 9000-AG	K 9010-AG	K 9020-AG
Abutment screw	K 60	K 61	K 62
Laboratory screw	K 60 L	K 61 L	K 62 L

Please note: This is used with Scanbody 2. generation.

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection	NP 3,5	RP 4,1	WP 5,1
Article No.	K 1400	K 1410	K 1420

Please note: The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 35 Ncm



Implant connection	NP 3,5	NP 3,5	NP 3,5
Gingiva height	2,5 mm	3,5 mm	4,5 mm
Article No.	K 4000	K 4100	K 4200

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 35 Ncm



Implant connection	RP 4,1	RP 4,1	RP 4,1
Gingiva height	2,5 mm	3,5 mm	4,5 mm
Article No.	K 4010	K 4110	K 4210

MedentiBASE® prosthetic components

- K 4840 MedentiBASE® titanium base incl. bridge screw
- K 7600 MedentiBASE® bridge screw UG
- Recommended torque: 15 Ncm



Material	Titanium grade 5 KV	Titanium grade 5 KV
Article No.	K 4840	K 7600

MedentiBASE® adhesive cap for passive-fit

- K 4800 MedentiBASE® adhesive cap Short incl. bridge screw
- K 4820 MedentiBASE® adhesive cap Long incl. bridge screw



Version	Short	Long
Article No.	K 4800	K 4820

MedentiBASE® scanbody

- Titanium grade 5 KV specially coated
- incl. bridge screw



Article No.	K 4900
-------------	--------

Please note: The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® accessories


- M 11-6 Placement instrument
MedentiBASE® abutment



Article No.

M 11-6

»» L-SERIES ««




The L-Series abutments are compatible with



Straumann® Bone Level®


DETAILED VIEW

Exact view of the direction of the angulation (indexing) of all angled abutments.



Type 1

angled over flat




Type 2

angled over edge

Laboratory implant

- Titanium grade 5 KV



Implant connection

NC 3,3

RC 4,1/4,8

Article No.

L 50

L 51

Abutment screw

- Titanium grade 5 KV
- Torx T6



Article No.

L 60

L 63

Please note:

The correct choice of the abutment screw is stated with the respective abutment in the catalogue.

Laboratory screw

- red coated



Article No.


L 60 L

Please note:

The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection

NC 3,3

RC 4,1/4,8

Chimney height

3,5 mm

3,5 mm

Gingiva height

1,0 mm

0,8 mm

Article No.

L 1000

L 1010

Abutment screw

L 60

L 60

Laboratory screw

L 60 L

L 60 L

Please note:

This is used with Scanbody 2. generation.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection	NC 3,3	RC 4,1/4,8
Chimney height	5.5 mm	5.5 mm
Gingiva height	1,0 mm	0,8 mm
Article No.	L 1100	L 1110
Abutment screw	L 60	L 60
Laboratory screw	L 60 L	L 60 L

Please note: This is used with Scanbody 2. generation.

Titanium base Bridges/ bars 2nd Generation

- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm




Implant connection	NC 3,3	RC 4,1/4,8
Chimney height	3.5 mm	3.5 mm
Gingiva height	1,0 mm	0,8 mm
Article No.	L 1500	L 1510
Abutment screw	L 60	L 63
Laboratory screw	L 60 L	

Please note: This is used with Scanbody 2. generation.

TI-Forms®-Abutment for Ceramill

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection	NC 3,3	RC 4,1/4,8
Diameter	11.5 mm	11.5 mm
Article No.	L 9000-AG	L 9010-AG
Abutment screw	L 60	L 60
Laboratory screw	L 60 L	L 60 L

Please note: This is used with Scanbody 2. generation.

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection	NC 3,3	RC 4,1/4,8
Article No.	L 1400	L 1410

Please note: The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 35 Ncm



Implant connection	NC 3,3	NC 3,3	NC 3,3	NC 3,3	NC 3,3
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	L 4000	L 4100	L 4200	L 4300	L 4400

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 35 Ncm



Implant connection	RC 4,1/4,8	RC 4,1/4,8	RC 4,1/4,8	RC 4,1/4,8	RC 4,1/4,8
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	L 4010	L 4110	L 4210	L 4310	L 4410

MedentiBASE® prosthetic components

- L 4840 MedentiBASE® titanium base incl. bridge screw
- L 7600 MedentiBASE® bridge screw Torx T6
- Recommended torque: 15 Ncm



Material	Titanium grade 5 KV	Titanium grade 5 KV
Article No.	L 4840	L 7600

MedentiBASE® adhesive cap for passive-fit

- L 4800 MedentiBASE® adhesive cap Short incl. bridge screw
- L 4820 MedentiBASE® adhesive cap Long incl. bridge screw



Version	Short	Long
Article No.	L 4800	L 4820

MedentiBASE® scanbody

- Titanium grade 5 KV specially coated
- incl. bridge screw



Article No.	L 4900
-------------	--------

Please note: The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® accessories

- M 11-6 Placement instrument
MedentiBASE® abutment



Article No.

M 11-6

» N-SERIES «




The N-Series abutments are compatible with



Straumann® Tissue Level®


DETAILED VIEW

Exact view of the direction of the angulation (indexing) of all angled abutments.



Type 1

angled over flat



Type 2

angled over corner

Laboratory implant

- Titanium grade 5 KV



Implant connection	NNC 3,5	RN 4,8	WN 6,5
Article No.	N 53	N 51	N 52

Abutment screw

- Titanium grade 5 KV
- Torx T6



Article No.	N 62	N 65
-------------	------	------

Please note: The correct choice of the abutment screw is stated with the respective abutment in the catalogue.

Laboratory screw

- red coated



Article No.	N 62 L	N 65 L
-------------	--------	--------

Please note: The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection	NNC 3,5	RN 4,8	WN 6,5
Chimney height	3.5 mm	3.5 mm	3.5 mm
Gingiva height	0,6 mm	0,1 mm	0,1 mm
Article No.	N 1030	N 1010	N 1020
Abutment screw	N 65	N 62	N 62
Laboratory screw	N 65 L	N 62 L	N 62 L

Please note: This is used with Scanbody 2. generation.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection	NNC 3,5	RN 4,8	WN 6,5
Chimney height	5.5 mm	5.5 mm	5.5 mm
Gingiva height	0,6 mm	0,1 mm	0,1 mm
Article No.	N 1130	N 1110	N 1120
Abutment screw	N 65	N 62	N 62
Laboratory screw	N 65 L	N 62 L	N 62 L

Please note: This is used with Scanbody 2. generation.

Titanium base Bridges/ bars 2nd Generation

- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection	NNC 3,5	RN 4,8	WN 6,5
Chimney height	3.5 mm	3.5 mm	3.5 mm
Gingiva height	0,35 mm	0,1 mm	0,1 mm
Article No.	N 1530	N 1510	N 1520
Abutment screw	N 65	N 62	N 62
Laboratory screw	N 65 L	N 62 L	N 62 L

Please note: This is used with Scanbody 2. generation.

TI-Forms®-Abutment for Ceramill

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 35 Ncm



Implant connection	NNC 3,5	RN 4,8	WN 6,5
Diameter	11.5 mm	11.5 mm	11.5 mm
Article No.	N 9030-AG	N 9010-AG	N 9020-AG
Abutment screw	N 65	N 62	N 62
Laboratory screw	N 65 L	N 62 L	N 62 L

Please note: This is used with Scanbody 2. generation.

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection	NNC 3,5	RN 4,8	WN 6,5
Article No.	N 1430	N 1410	N 1420

Please note: The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 35 Ncm



Implant connection	NNC 3,5	NNC 3,5	NNC 3,5	NNC 3,5	NNC 3,5
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	N 4030	N 4130	N 4230	N 4330	N 4430

MedentiBASE® prosthetic components

- NNC 4840 MedentiBASE® titanium base incl. bridge screw
- NNC 7600 MedentiBASE® bridge screw Torx T6
- Recommended torque: 15 Ncm



Material	Titanium grade 5 KV	Titanium grade 5 KV
Article No.	NNC 4840	NNC 7600

MedentiBASE® adhesive cap for passive-fit

- NNC 4800 MedentiBASE® adhesive cap Short incl. bridge screw
- NNC 4820 MedentiBASE® adhesive cap Long incl. bridge screw



Version	Short	Long
Article No.	NNC 4800	NNC 4820

MedentiBASE® scanbody

- Titanium grade 5 KV specially coated
- incl. bridge screw
- sterilisable and for intra-oral scanning



Article No.	NNC 4900
-------------	----------

Please note: The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® accessories

- M 11-6 Placement instrument MedentiBASE® abutment



Article No.	M 11-6
-------------	--------

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 35 Ncm



Implant connection	RN 4,8
Gingiva height	0 mm
Article No.	N 4010

MedentiBASE® prosthetic components

- N 7600 MedentiBASE® bridge screw Torx T6
- Recommended torque: 15 Ncm



Material	Titanium grade 5 KV
Article No.	N 7600

MedentiBASE® adhesive cap for passive-fit

- N 4800 MedentiBASE adhesive cap Short incl. bridge screw



Version	Short
Article No.	N 4800

MedentiBASE® scanbody

- Titanium grade 5 KV specially coated
- incl. bridge screw



Article No.	N 4900
-------------	--------

Please note: The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® accessories

- M 11-6 Placement instrument MedentiBASE® abutment
- N 51 Laboratory implant



Article No.	M 11-6	N 51
-------------	--------	------

» R-SERIES «




The R-Series abutments are compatible with



- Zimmer Dental®** - Tapered Screw-Vent®
MIS® - SEVEN® Internal Hex
BioHorizons® - Tapered Internal
- Internal Plus
- Tapered Tissue Level


DETAILED VIEW

Exact view of the direction of the angulation (indexing) of all angled abutments.



Type 1

angled over flat



Type 2

angled over corner












Comparison chart




Zimmer Dental®

MIS®

BioHorizons®

Zimmer Dental®	Platform 3,5 mm		Platform 4,5 mm	Platform 5,7 mm
Implant Ø	3,70	4,10	4,70	6,00
Laboratory implant Ø	3,50	3,50	4,50	5,70
Article Number	IA3	IA3	IA4	IA5
Implant connection				
Medentika R-Series Laboratory implant	3,5	3,5	4,5	5,7
Article Number	R 50	R 50	R 51	R 52

MIS®	Narrow- Platform	Standard- Platform			Wide- Platform	
Implant Ø	3,30	3,30	3,75	4,20	5,00	6,00
Laboratory implant Ø	3,30	4,20	4,20	4,20	4,50	4,50
Article Number	MN-RSM10	MD-RSM10	MD-RSM10	MD-RSM10	MW-RSM10	MW-RSM10
Implant connection						
Medentika R-Series Laboratory implant	XX	3,50	3,50	3,50	4,50	4,50
Article Number	XX	R 50	R 50	R 50	R 51	R 51

BioHorizons®	Platform 3,0 mm	Platform 3,5 mm	Platform 4,5 mm	Platform 5,7 mm
Implant Ø	3,00	3,80	4,60	5,80
Laboratory implant Ø	3,00	3,50	4,50	5,70
Article Number	TP3IA	SYIA	SGIA	SBIA
Implant connection				
Medentika R-Series Laboratory implant	XX	3,50	4,50	5,70
Article Number	XX	R 50	R 51	R 52

Laboratory implant

- Titanium grade 5 KV



Implant connection

D 3,5

D 4,5

D 5,7

Article No.

R 50

R 51

R 52

Abutment screw

- Titanium grade 5 KV
- Hex 1.26



Article No.

R 60

Please note:

The correct choice of the abutment screw is stated with the respective abutment in the catalogue.

Laboratory screw

- red coated



Article No.

R 60 L

Please note:

The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 30 Ncm



Implant connection

D 3,5

D 4,5

D 5,7

Chimney height

3.5 mm

3.5 mm

3.5 mm

Gingiva height

0,3 mm

0,3 mm

0,3 mm

Article No.

R 1000

R 1010

R 1020

Abutment screw

R 60

R 60

R 60

Laboratory screw

R 60 L

R 60 L

R 60 L

Please note:

This is used with Scanbody 2. generation.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 30 Ncm



Implant connection	D 3,5	D 4,5	D 5,7
Chimney height	5.5 mm	5.5 mm	5.5 mm
Gingiva height	0,3 mm	0,3 mm	0,3 mm
Article No.	R 1100	R 1110	R 1120
Abutment screw	R 60	R 60	R 60
Laboratory screw	R 60 L	R 60 L	R 60 L

Please note: This is used with Scanbody 2. generation.

Titanium base Bridges/ bars 2nd Generation

- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 30 Ncm



Implant connection	D 3,5	D 4,5	D 5,7
Chimney height	3.5 mm	3.5 mm	3.5 mm
Gingiva height	0,3 mm	0,3 mm	0,3 mm
Article No.	R 1500	R 1510	R 1520
Abutment screw	R 60	R 60	R 60
Laboratory screw	R 60 L	R 60 L	R 60 L

Please note: This is used with Scanbody 2. generation.

TI-Forms®-Abutment for Ceramill

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 30 Ncm



Implant connection	D 3,5	D 4,5	D 5,7
Diameter	11.5 mm	11.5 mm	11.5 mm
Article No.	R 9000-AG	R 9010-AG	R 9020-AG
Abutment screw	R 60	R 60	R 60
Laboratory screw	R 60 L	R 60 L	R 60 L

Please note: This is used with Scanbody 2. generation.

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection	D 3,5	D 4,5	D 5,7
Article No.	R 1400	R 1410	R 1420

Please note: The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 30 Ncm



Implant connection	D 3,5	D 3,5	D 3,5	D 3,5	D 3,5
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	R 4000	R 4100	R 4200	R 4300	R 4400

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 30 Ncm



Implant connection	D 4,5	D 4,5	D 4,5	D 4,5	D 4,5
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	R 4010	R 4110	R 4210	R 4310	R 4410

MedentiBASE® prosthetic components

- R 4840 MedentiBASE® titanium base incl. bridge screw
- R 7600 MedentiBASE® bridge screw Hex 1.26
- Recommended torque: 15 Ncm



Material	Titanium grade 5 KV	Titanium grade 5 KV
Article No.	R 4840	R 7600

MedentiBASE® adhesive cap for passive-fit

- R 4800 MedentiBASE® adhesive cap Short incl. bridge screw
- R 4820 MedentiBASE® adhesive cap Long incl. bridge screw



Version	Short	Long
Article No.	R 4800	R 4820

MedentiBASE® scanbody

- Titanium grade 5 KV specially coated
- incl. bridge screw



Article No.	R 4900
-------------	--------

Please note: The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® accessories

- M 11-6 Placement instrument
MedentiBASE® abutment



Article No.

M 11-6

» S-SERIES «



The S-Series abutments are compatible with



DENTSPLY Implants®
ASTRA TECH OsseoSpeed® TX

Laboratory implant

- Titanium grade 5 KV



Implant connection	D 3,0	D 3,5/4,0	D 4,5/5,0
Article No.	S 53	S 50	S 52

Abutment screw

- Titanium grade 5 KV
- Hex 1.26



Article No.	S 60	S 61	S 67
-------------	------	------	------

Please note: The correct choice of the abutment screw is stated with the respective abutment in the catalogue.

Laboratory screw

- red coated



Article No.	S 60 L	S 61 L	S 67 L
-------------	--------	--------	--------

Please note: The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque:
15 Ncm: D 3,0
25 Ncm: D 3,5/4,0
25 Ncm: D 4,5/5,0



Implant connection	D 3,0	D 3,5/4,0	D 3,5/4,0	D 4,5/5,0	D 4,5/5,0
Chimney height	3,5 mm	3,5 mm	3,5 mm	3,5 mm	3,5 mm
Gingiva height	0,6 mm	0,6 mm	1,1 mm	0,6 mm	1,1 mm
Article No.	S 1030	S 1000	S 1200	S 1020	S 1220
Abutment screw	S 67	S 60	S 60	S 61	S 61
Laboratory screw	S 67 L	S 60 L	S 60 L	S 61 L	S 61 L

Please note: This is used with Scanbody 2. generation.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 15 Ncm: D 3,0
25 Ncm: D 3,5/4,0
25 Ncm: D 4,5/5,0




Implant connection	D 3,0	D 3,5/4,0	D 3,5/4,0	D 4,5/5,0	D 4,5/5,0
Chimney height	5,5 mm	5,5 mm	5,5 mm	5,5 mm	5,5 mm
Gingiva height	0,6 mm	0,6 mm	1,1 mm	0,6 mm	1,1 mm
Article No.	S 1130	S 1100	S 1300	S 1120	S 1320
Abutment screw	S 67	S 60	S 60	S 61	S 61
Laboratory screw	S 67 L	S 60 L	S 60 L	S 61 L	S 61 L

Please note: This is used with Scanbody 2. generation.

Titanium base Bridges/ bars 2nd Generation

- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm




Implant connection	D 3,5/4,0	D 4,5/5,0
Chimney height	3,5 mm	3,5 mm
Gingiva height	0,6 mm	0,6 mm
Article No.	S 1500	S 1520
Abutment screw	S 60	S 61
Laboratory screw	S 60 L	S 61 L

Please note: This is used with Scanbody 2. generation.

TI-Forms®-Abutment for Ceramill

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm



Implant connection	D 3,5/4,0	D 4,5/5,0
Diameter	11,5 mm	11,5 mm
Article No.	S 9000-AG	S 9020-AG
Abutment screw	S 60	S 61
Laboratory screw	S 60 L	S 61 L

Please note: This is used with Scanbody 2. generation.

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection	D 3,0	D 3,5/4,0	D 4,5/5,0
Article No.	S 1430	S 1400	S 1420

Please note: The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 25 Ncm



Implant connection	D 3,5/4,0	D 3,5/4,0	D 3,5/4,0	D 3,5/4,0	D 3,5/4,0
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	S 4000	S 4100	S 4200	S 4300	S 4400

MedentiBASE® abutment straight


- Titanium grade 5 KV
- Recommended torque: 25 Ncm



Implant connection	D 4,5/5,0	D 4,5/5,0	D 4,5/5,0	D 4,5/5,0	D 4,5/5,0
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	S 4020	S 4120	S 4220	S 4320	S 4420

MedentiBASE® prosthetic components

- S 4840 MedentiBASE® titanium base incl. bridge screw
- S 7600 MedentiBASE® bridge screw Hex 1.26
- Recommended torque: 15 Ncm



Material	Titanium grade 5 KV	Titanium grade 5 KV
Article No.	S 4840	S 7600

MedentiBASE® adhesive cap for passive-fit

- S 4800 MedentiBASE® adhesive cap Short incl. bridge screw
- S 4820 MedentiBASE® adhesive cap Long incl. bridge screw



Version	Short	Long
Article No.	S 4800	S 4820

MedentiBASE® scanbody

- Titanium grade 5 KV specially coated
- incl. bridge screw



Article No.	S 4900
-------------	--------

Please note: The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® accessories

- M 11-6 Placement instrument
MedentiBASE® abutment



Article No.

M 11-6

» T-SERIES «




The T-Series abutments are compatible with



DENTSPLY Implants® XiVE® S


DETAILED VIEW

Exact view of the direction of the angulation (indexing) of all angled abutments.



Type 1

angled over flat



Type 2

angled over corner

Laboratory implant

- Stainless steel



Implant connection	D 3,4	D 3,8	D 4,5	D 5,5
Article No.	T 50	T 55	T 51	T 52

Abutment screw

- Titanium grade 5 KV
- Hex 1.2



Article No.	T 60
-------------	------

Please note: The correct choice of the abutment screw is stated with the respective abutment in the catalogue.

Laboratory screw

- red coated



Article No.	T 60 L
-------------	--------

Please note: The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm



Implant connection	D 3,4	D 3,8	D 4,5	D 5,5
Chimney height	3,5 mm	3,5 mm	3,5 mm	3,5 mm
Gingiva height	0,3 mm	0,3 mm	0,6 mm	0,6 mm
Article No.	T 1000	T 1005	T 1010	T 1020
Abutment screw	T 60	T 60	T 60	T 60
Laboratory screw	T 60 L	T 60 L	T 60 L	T 60 L

Please note: This is used with Scanbody 2. generation.

Titanium base Zirconium abutment 2nd Generation

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm



Implant connection	D 3,4	D 3,8	D 4,5	D 5,5
Chimney height	5.5 mm	5.5 mm	5.5 mm	5.5 mm
Gingiva height	0,3 mm	0,3 mm	0,6 mm	0,6 mm
Article No.	T 1100	T 1105	T 1110	T 1120
Abutment screw	T 60	T 60	T 60	T 60
Laboratory screw	T 60 L	T 60 L	T 60 L	T 60 L

Please note: This is used with Scanbody 2. generation.

Titanium base Bridges/ bars 2nd Generation

- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm



Implant connection	D 3,4	D 3,8	D 4,5	D 5,5
Chimney height	3.5 mm	3.5 mm	3.5 mm	3.5 mm
Gingiva height	0,3 mm	0,3 mm	0,6 mm	0,6 mm
Article No.	T 1500	T 1505	T 1510	T 1520
Abutment screw	T 60	T 60	T 60	T 60
Laboratory screw	T 60 L	T 60 L	T 60 L	T 60 L

Please note: This is used with Scanbody 2. generation.

TI-Forms®-Abutment for Ceramill

- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm



Implant connection	D 3,4	D 3,8	D 4,5	D 5,5
Diameter	11.5 mm	11.5 mm	11.5 mm	11.5 mm
Article No.	T 9000-AG	T 9005-AG	T 9010-AG	T 9020-AG
Abutment screw	T 60	T 60	T 60	T 60
Laboratory screw	T 60 L	T 60 L	T 60 L	T 60 L

Please note: This is used with Scanbody 2. generation.

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection	D 3,4	D 3,8	D 4,5/5,5
Article No.	T 1400	T 1405	T 1410

Please note: The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 25 Ncm



Implant connection	D 3,4	D 3,4	D 3,4	D 3,4	D 3,4
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	T 4000	T 4100	T 4200	T 4300	T 4400

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 25 Ncm



Implant connection	D 3,8	D 3,8	D 3,8	D 3,8	D 3,8
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	T 4005	T 4105	T 4205	T 4305	T 4405

MedentiBASE® abutment straight

- Titanium grade 5 KV
- Recommended torque: 25 Ncm



Implant connection	D 4,5	D 4,5	D 4,5	D 4,5	D 4,5
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	T 4010	T 4110	T 4210	T 4310	T 4410

MedentiBASE® prosthetic components

- T 4840 MedentiBASE® titanium base incl. bridge screw
- T 7600 MedentiBASE® bridge screw Hex 1.2
- Recommended torque: 15 Ncm



Material	Titanium grade 5 KV	Titanium grade 5 KV
Article No.	T 4840	T 7600

MedentiBASE® adhesive cap for passive-fit

- T 4800 MedentiBASE® adhesive cap Short incl. bridge screw
- T 4820 MedentiBASE® adhesive cap Long incl. bridge screw



Version	Short	Long
Article No.	T 4800	T 4820

MedentiBASE® scanbody

- Titanium grade 5 KV specially coated
- incl. bridge screw



Article No.

T 4900

Please note: The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® accessories

- M 11-6 Placement instrument
MedentiBASE® abutment



Article No.

M 11-6

»» Y-SERIES ««



The Y-Series abutments are compatible with



DENTSPLY Implants®
ANKYLOS®

Laboratory implant

- Stainless steel



Implant connection

D 3,5-7,0

Article No.

Y 50

Abutment screw

- Titanium grade 5 KV
- Hex 1.0



Article No.

Y 61

Please note:

The correct choice of the abutment screw is stated with the respective abutment in the catalogue.

Laboratory screw

- red coated



Article No.

Y 61 L

Please note:

The correct choice of the laboratory screw is stated with the respective abutment in the catalogue.

Titanium base Zirconium abutment 2nd Generation

- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm



Implant connection

D 3,5-7,0

D 3,5-7,0

Chimney height

3,5 mm

3,5 mm

Gingiva height

0,65 mm

1,15 mm

Article No.

Y 1500

Y 1700

Abutment screw

Y 61

Y 61

Laboratory screw

Y 61 L

Y 61 L

Please note:

This is used with Scanbody 2. generation.

Titanium base Zirconium abutment 2nd Generation

- rotating
- Titanium grade 5 KV
- incl. abutment screw
- Recommended torque: 25 Ncm



Implant connection	D 3,5-7,0	D 3,5-7,0
Chimney height	5.5 mm	5.5 mm
Gingiva height	0,65 mm	1,15 mm
Article No.	Y 1600	Y 1800
Abutment screw	Y 61	Y 61
Laboratory screw	Y 61 L	Y 61 L

Scanbody 2nd Generation

- Titanium grade 5 KV
- specially coated
- incl. retention screw



Implant connection	D 3,5-7,0
Article No.	Y 1400

Please note: The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE abutment straight

- Titanium grade 5 KV
- Recommended torque: 25 Ncm



Implant connection	D 3,5-7,0	D 3,5-7,0	D 3,5-7,0	D 3,5-7,0	D 3,5-7,0
Gingiva height	0,5 mm	1,5 mm	2,5 mm	3,5 mm	4,5 mm
Article No.	Y 4000	Y 4100	Y 4200	Y 4300	Y 4400

MedentiBASE prosthetic components

- Y 4840 MedentiBASE® titanium base incl. bridge screw
- Y 7600 MedentiBASE® bridge screw Hex 1.0
- Recommended torque: 15 Ncm



Material	Titanium grade 5 KV	Titanium grade 5 KV
Article No.	Y 4840	Y 7600

MedentiBASE adhesive cap for passive-fit

- Y 4800 MedentiBASE® adhesive cap Short incl. bridge screw
- Y 4820 MedentiBASE® adhesive cap Long incl. bridge screw



Version	Short	Long
Article No.	Y 4800	Y 4820

MedentiBASE® scanbody

- Titanium grade 5 KV specially coated
- incl. bridge screw



Article No.

Y 4900

Please note: The Scanbody is sterilisable and for intra-oral scanning.

MedentiBASE® accessories

- M 11-6 Placement instrument
MedentiBASE® abutment



Article No.

M 11-6

»» Tools ««





Placement instrument MedentiBASE® abutment

- Hardened stainless steel



Version	Manual and ratchet
Type	2-piece
Article No.	M 11-6

Torque ratchet

- with infinitely variable torque setting
- up to max. 40 Ncm
- Hardened stainless steel



Article No.	M 12
-------------	------

Wooden tool box Placement instrument Manual and ratchet



Version	long
Article No.	MB 04K
Content:	Hex 1,26 / Hex 1,2 / Hex 1,0 / Torx T6 / UG

Compartment box Abutments and scanbodies

- 30 compartments
- Polystyrene



Version	without contents
Article No.	MB 04

HERSTELLER:

MEDENTIKA GMBH

Hammerweg 8-10
76549 Hügelsheim
Fon +49 7229 69912-0
Fax +49 7229 69912-20
info@medentika.de
www.medentika.de

VERTRIEB:

AUSTRIA (HEADQUARTERS)

Amann Girschbach AG
Koblach, Austria
Fon +43 5523 62333-105
austria@amanngirschbach.com

GERMANY

Amann Girschbach GmbH
Pforzheim, Germany
Fon +49 7231 957-100
germany@amanngirschbach.com

NORTH AMERICA

Amann Girschbach North America, LP
Charlotte, NC, U.S.A.
Fon +1 704 837 1404
america@amanngirschbach.com

BRASIL

Amann Girschbach Brasil LTDA
Curitiba, Brasil
Fon +55 41 3287 0897
brasil@amanngirschbach.com

ASIA

Amann Girschbach Asia PTE LTD.
Singapore, Asia
Fon +65 6592 5190
singapore@amanngirschbach.com

CHINA

Amann Girschbach China Co., Ltd.
Beijing, China
Fon +86 10 8886 6064
china@amanngirschbach.com